

Human Factor XIII Matched Pair Antibodies for EIA (5 x 96 Tests)

Ref#: F13-EIA

Store at -10 to -20°C

For Research Use Only.

Not for Use in Diagnostic Procedures.

For *in vitro* use only.

INTENDED USE:

Human Factor XIII Matched Pair Antibodies for EIA are intended for use with in-house enzyme-linked immunosorbent assays for measuring human Factor XIII in plasma, or in any biological fluid where human FXIII can be present. The results obtained should be for research purposes only and not used for patient diagnosis or treatment.

SUMMARY:

Coagulation Factor XIII (FXIII) is a non-covalent tertrameric complex (320 kDA) of two A-subunits and two B-subunits. The B subunits (76 kDa each) act as a carrier protein for the A subunits (83 kDa each). The normal Factor XIII protease zymogen concentration in human plasma is about 25 μ g/mL (80 nM).

ASSAY PRINCIPLE:

The diluted plasma sample or biological fluid is introduced into one of the microwells of a micro ELISA plate which has been pre-coated with anti-human FXIII antibody. When present in the added material, FXIII binds to the anti-human polyclonal antibody. Following a washing step, the remaining bound antibodies are revealed with a detection antibody, anti-human peroxidase conjugate, which reacts specifically with human FXIII. Following another washing step, the peroxidase substrate, o-Phenylenediamine (OPD) in presence of hydrogen peroxide (H₂O₂), is introduced and a yellow color develops. The color turns orange when the reaction is stopped with sulfuric acid. The color developed is directly proportional to the amount of FXIII present in the tested sample.

REAGENTS:

Required Materials provided (enough for 5x96 Tests):

<u>C:</u> Capture Antibody (F13-EIA-C). 1 vial of 0.5 mL polyclonal affinity purified antibody specific for human FXIII. For coating plates. Supplied in a 50% v/v glycerol solution. Yellow cap.
<u>D:</u> Detecting Antibody (F13-EIA-D). 1 vial of 0.5 mL polyclonal antibody

D: Detecting Antibody (F13-EIA-D). 1 vial of 0.5 mL polyclonal antibody specific for human FXIII, coupled to peroxidase. For detecting captured FXIII. Supplied in a 50% v/v glycerol solution. Red cap.

 $\underline{\textbf{Note}}:$ Antibodies are provided in a glycerol solution (50% v/v) and should be stored at -10 to -20°C. Vials should be tightly capped. Do not store in frost-free freezers.

Antibodies can be centrifuged briefly in a micro-centrifuge to gather residual reagent from the cap and walls of the tube.

In their original packaging, before use, when stored at -10 to -20 $^{\circ}\text{C}$, the unopened antibodies are stable until the expiration date printed on the vial.

Required Materials not provided:

Optimum performance can be obtained when the following solutions and assay conditions are used.

- Micro ELISA plates with hydrophilic surface designed for high binding of IgG. For example, 96-well Immulon 4-HBX.
- Coating Solution. 50 mM Carbonate. Dissolve 1.59 g of Na2Cl3 and 2.93 g of NaHCO3 in distilled water to a final volume of 1 L and adjust pH to 9.6. Store at 2-8°C for 1 month.

- **Phosphate-Buffered Saline (PBS).** [For preparation of wash and blocking solutions.] Dissolve 8.0 g NaCl, 1.15 g Na₂HPO₄ 0.2 g KH₂PO₄ and 0.2 g KCl in distilled water to a final volume of 1 L and adjust **pH to 7.4**. Store up to 1 month at 2-8°C, discard if there is evidence of microbial growth.
- Sample Diluent and Wash Solution: PBS/Tween-20 (0.1% v/v). Add 1.0 mL of Tween-20 to 1 L of PBS and adjust pH to 7.4. Store at 2-8°C up to 1 week.
- Conjugate Diluent: HBS/BSA/T-20. Dissolve 5.95 g HEPES (free acid), 1.46 g NaCl and 2.5 g of Bovine Serum Albumin (Sigma-RIA grade) in 200 mL of H₂O. Add 0.25 mL Tween-20 and adjust pH to 7.2 with NaOH; add H₂O to a final volume of 250 mL. Aliquot and store frozen at -20°C.
- **Blocking Solution:** PBS/BSA (1% w/v). Dissolve 2.5 g of Bovine Serum Albumin (Sigma-RIA grade) in 200 mL of PBS and adjust **pH to 7.4**; add PBS to final volume of 250 mL. Aliquot and store frozen at -20°C.
- Substrate Solution: Citrate-Phosphate buffer. Dissolve 2.6 g Citric Acid
 and 6.9 g Na₂HPO₄ in distilled water up to a final volume of 500 mL and
 adjust pH to 5.0. Store at 2-8°C up to 1 month.
- OPD Substrate: o-phenylenediamine.2HCl Toxic! (5 mg tablets: Sigma #P-6912). Prepare immediately before use. Dissolve 5 mg OPD in 12 mL Substrate Solution and then add 12 µL 30% H₂O₂. Do not store.
 Stop Solution: 2.5 M H₂SO₄. Corrosive! Generates heat on dilution!
- Stop Solution: 2.5 M H₂SO₄. Corrosive! Generates heat on dilution!
 Handle with great care. Avoid any skin and eye contact. Wear protective glasses and gloves when handling. Carefully add 13.9 mL 18 M H₂SO₄ to 86 mL distilled H₂O. Store at room temperature.
- Reference standards for Factor XIII which have the same matrix and anticoagulant as the samples to be tested.
- Micro ELISA plate washing equipment and shaker.
- Micro ELISA plate reader with a wavelength set up at 490 nm.

PROCEDURE

- 1. Coat ELISA plate: Dilute the Capture Antibody with Coating Solution 1/100 (use polypropylene tube) and immediately add 100 μL to every well in the plate. Incubate for 2 hours @ 22°C or overnight @ 2-8°C.
- **2. Blocking:** Empty contents of plate and add 150 μ L of Blocking Solution to every well and incubate for 60 minutes @ 22°C. This step blocks any remaining binding sites on the plastic wells. Wash plate 3X with Wash Solution.
- **3. Samples:** Dilute FXIII Reference standard with Sample Diluent 1/200 (100%) then serially dilute by halves down to 1/6400 (3.13%). Dilute sample plasmas or biological fluid with Sample Diluent 1/400, 1/800 and 1/1600. Apply 100 μL per well and incubate plate @ 22°C for 60 minutes. Wash plate 3X with Wash Solution. [Plasma samples should not be applied at dilutions lower than 1/20, as falsely high readings may result.]
- 4. Detecting Antibody: Dilute the Detecting Antibody with Conjugate Diluent 1/100 and apply 100 μL to each well. Incubate plate @ 22°C for 60 minutes. Wash plate 3X with Wash Solution.
- **5. OPD Substrate:** Apply 100 μ L of freshly prepared OPD substrate to each well. Allow color to develop for 10-15 minutes then stop color reaction with the addition of 50 μ L per well of Stop Solution. Read the plate at a wavelength of **490 nm.** [Optimal color development time is the time required to obtain A490 \geq 1.000 for the 100% reference point, not to exceed 20 minutes.]

Additional Notes:

- Do not allow the wells to become dry. Keep plate covered or in a humid chamber during incubations.
- Rheumatoid factor in samples may bind to the capture and/or detecting antibodies and cause interference in the ELISA assay.
- **6. Calibration Curve:** On bi-logarithmic graph paper, plot the known Factor XIII concentrations on abscissa and the corresponding absorbance (A490) on ordinates in order to establish the calibration curve.

RESULTS:

From the constructed calibration curve, directly determine the Factor XIII concentration and multiply by the appropriate dilution factor.

CoaChrom Diagnostica GmbH

Hauptstrasse 5, Å-2344 Maria Enzersdorf, Austria Tel: +43-1-236 2221 | Deutschland: 0800-24 66 330 Fax: +43-1-236 22211 | Deutschland: 0800-24 66 333 Email: info@coachrom.com

www.coachrom.com