

****REPRESENTATIVE DATASHEET****

**Matched-Pair Antibody Set
for ELISA of murine
Antithrombin antigen (MATIII)**

Sufficient reagent for 4 x 96 well plates

Product #: MATIII-EIA
Lot #: SAMPLE
Expiry Date: SAMPLE

Store at -10 to -20°C

For Research Use Only
Not for use in diagnostic procedures.

Description of Antithrombin (ATIII)

Antithrombin, also known as Antithrombin III (ATIII), is a member of the SERPIN family of proteinase inhibitors and the primary inhibitor of thrombin in plasma. It is produced in the liver and circulates in human plasma at ~200 µg/ml (~3.5 µM). Antithrombin inhibits a broad spectrum of serine proteases including thrombin, activated forms of factor X, factor IX, factor XI, factor XII, as well as kallikrein, plasmin and urokinase. Enzyme inhibition by antithrombin occurs through proteolytic cleavage at Arg³⁸⁵-Ser³⁸⁶ and subsequent rapid formation of a stable, inactive 1:1 enzyme-antithrombin complex. Heparin has a profound accelerating effect on the inhibitory activity of antithrombin towards some enzymes. For example, the rate of inhibition of thrombin and activated factor X is increased 1000-fold in the presence of optimal concentrations of heparin, whereas heparin has relatively little effect on the inhibition rate of activated factor XI, activated factor XII and kallikrein. Human antithrombin is a single chain molecule with a molecular weight of 59 kDa. Interaction with thrombin results in an SDS-stable thrombin-antithrombin complex of 96 kDa¹⁻³.

Principle of Sandwich-style ELISA

Affinity-purified antibody to ATIII is coated onto the wells of a microtitre plate. The plates are washed and plasma or other fluids containing MATIII are applied. The coated antibody will capture the MATIII in the sample. After washing the plate to remove unbound material, a peroxidase conjugated second antibody to ATIII is added to the plate to bind to the captured MATIII. After washing the plate to remove unbound conjugated antibody, the peroxidase activity is expressed by incubation with o-phenylenediamine (OPD). After a fixed development time the reaction is quenched with the addition of H₂SO₄ and the colour produced is quantified using a microplate reader. The colour generated is proportional to the concentration of MATIII present in the sample.

Supplied Materials:

1. Capture Antibody (MATIII-EIA-C): One yellow-capped vial containing 0.4 ml of polyclonal affinity purified anti-ATIII antibody for coating plates.

2. Detecting Antibody (MATIII-EIA-D): One red-capped tube containing 0.4 ml of affinity purified, peroxidase conjugated polyclonal anti-ATIII antibody for detection of captured MATIII.

Note: Antibodies are supplied in a 50% (v/v) glycerol solution for storage at -10 to -20°C. Keep vials tightly capped. Do not store in frost-free freezers.

Materials Required but not Provided:

1. Coating Buffer: 50 mM Carbonate
1.59g of Na₂CO₃ and 2.93g of NaHCO₃ up to 1 litre. Adjust pH to 9.6. Store at 2-8°C up to 1 month.

2. PBS: (base for wash buffer and sample diluent)
8.0g NaCl, 1.15g Na₂HPO₄, 0.2g KH₂PO₄ and 0.2g KCl, up to 1 litre. Adjust pH to 7.4, if necessary. Store up to 1 month at 2-8°C, discard if there is evidence of microbial growth.

3. Wash Buffer and Sample Diluent: PBS-Tween (0.1%,v/v)
To 1 litre of PBS add 1.0 ml of Tween-20.
Check that the pH is 7.4. Store at 2-8°C up to 1 week.

4. Detecting Antibody Diluent: HBS-BSA-T20
5.95g HEPES (free acid), 1.46g NaCl, 2.5g Bovine Serum Albumin (Sigma, RIA grade) dissolved in 200 ml H₂O. Add 0.25 ml of Tween-20, check and adjust pH to 7.2 with NaOH, then make up to a final volume of 250 ml with H₂O. Aliquot and store frozen at -20°C.

5. Substrate Buffer: Citrate-Phosphate buffer pH 5.0
2.6g Citric acid and 6.9g Na₂HPO₄ up to a final volume of 500 ml with purified H₂O. Store at 2-8°C up to 1 month.

6. OPD Substrate: (o-Phenylenediamine.2HCl) Toxic!
(5mg tablets: Sigma # P-6912). Make up immediately before use. Dissolve 5mg OPD in 12 ml substrate buffer then add 12 µl 30% H₂O₂. Do not store.

7. Stopping Solution: 2.5 M H₂SO₄
Caution: VERY CORROSIVE! GENERATES HEAT ON DILUTION! Where stock sulphuric acid is 18 Molar, add 13.9 ml to 86 ml H₂O. Store at room temperature.

8. Other:
Microplates, 96-well Immulon 4-HBX (<http://www.labsystems.fi>)
Microplate washer (optional)
Microplate reader.

Assay Procedure:

1. Coating of plates:

Dilute the capture antibody 1/100 in coating buffer (preferably in a polypropylene tube) and immediately add 100 µl to every well in the plate. Incubate 2 hours at room temperature.

2. Blocking:

Blocking is not required under the conditions described. Washing the plate with PBS-Tween is sufficient to block non-specific interactions. Wash plate X 3 with wash buffer.

3. Samples:

Normal mouse plasma is diluted 1/4,000 (100%) then serial 1/2's down to 1/128,000 (3.13%). Sample plasmas are diluted 1/8,000, 1/16,000 & 1/32,000. All dilutions are made in PBS-Tween. Apply 100 µl/well and incubate plate @ 22°C for 60 minutes. Wash plate X 3 with wash buffer.

4. Detecting Antibody:

Dilute the detecting antibody 1/100 in HBS-BSA-T20 sample diluent and apply 100 µl to each well. Incubate plate @ 22°C for 60 minutes. Wash plate X 3 with wash buffer.

5. OPD Substrate:

Apply 100 µl of freshly prepared OPD substrate to every well. Allow colour to develop for 5-10 minutes then stop colour reaction with the addition of 50 µl/well of 2.5 M H₂SO₄. The plate can be read at wavelength of 490 nm.

Calculation of Results:

The construction of a proper reference curve is of no less importance than any other aspect of the assay. A reference curve should be constructed by plotting the known concentration of standards versus absorbance. This can be done manually using graph paper, or by using curve-fitting computer software. In our experience, the dose response curves of most immunoassays tend to be sigmoid in shape. Although linear regions can be identified within the curve, the best overall fit is often obtained using an algorithm that provides a weighted theoretical model of fit throughout the entire curve, such as a 4-parameter or 5-parameter logistic curve fit^{4,5}. In general, the simplest model that defines the concentration-response relationship should be used⁶.

The "back-fit" test is a simple and reliable method to determine if a curve-fitting method is appropriate. In this test, the apparent concentrations for the absorbance values of each standard point are read from the reference curve. The derived values are compared to the assigned values. An appropriate curve fitting method will produce derived values that closely match assigned values throughout the range of the curve, within user-defined limits⁶. The coefficient of determination (R^2) is a valuable indicator of the overall fit, but should not be used by itself in the selection of a curve fitting method, as a poor fit in a particular region of the curve may not be evident from this value alone^{5,6}.

In the quality control of this product we have determined that under the conditions described above, a reference curve that is constructed using serial dilutions of normal pooled plasma, will produce a correlation coefficient (R^2) of at least 0.980 using a log-log fit, and an R^2 of at least 0.990 using a 4-parameter logistic curve fit algorithm. However, the performance characteristics of in-house assays developed using this product in other laboratories may vary slightly from ours. Different curve fitting methods may be employed but we recommend that the back-fit test be applied as evidence that the fitting method is appropriate.

Technical Notes:

- This paired antibody product is intended to facilitate the end user in establishing an in-house immunoassay for research purposes only. It must not be used for diagnostic applications. Assay validation is the responsibility of the end user and should be done according to user-defined protocols⁶.
- Reference calibrators should be of the same matrix and anticoagulant as the samples to be tested (example serum or plasma, citrate or EDTA)
- Do not use samples diluted less than 1/100, as falsely high readings may result.
- The optimal colour development time should be determined empirically as the time required to obtain an absorbance of at least 1.000 at 490 nm for the 100% reference point, not to exceed 20 minutes.
- Rheumatoid factor in samples may interfere in ELISA by binding to the capture and/or detecting antibodies.
- The wells should not be allowed to become dry. Keep plate covered or in a humid chamber during incubations.
- Capture antibody is supplied in a 50% glycerol solution and can be centrifuged briefly in a micro-centrifuge to gather residual reagent from the cap and walls of the tube.

References:

1. Damus PS, Rosenberg RD; Antithrombin - Heparin Cofactor; Methods in Enzymology 45, pp 653-669, 1976.
2. Harpel PC; Blood Proteolytic Enzyme Inhibitors: Their Role in Modulating Blood Coagulation and Fibrinolytic Enzyme Pathways; in Hemostasis and Thrombosis, eds. RW Colman, J Hirsh, VJ Marder and EW Salzman, pp. 738-747, J.B. Lippincott Co., Philadelphia PA, USA, 1982.
3. Griffith MJ, Noyes CM, Church FC; Reactive Site Peptide Structural Similarity between Heparin Cofactor II and Antithrombin III; JBC:260, pp 2218-2225, 1985.
4. Nix, B, Wild D, in Immunoassays, A Practical Approach, editor J.P. Gosling, pp. 239-261, Oxford University Press, 2000.
5. NCCLS. Evaluation of the Linearity of Quantitative Analytical Methods; Proposed Guideline - Second Edition. NCCLS Document EP6-P2 (ISBN 1-56238-446-5, NCCLS, Wayne, Pennsylvania USA, 2001.
6. FDA Guidance for Industry. Bioanalytical Method Validation; May 2001, available on the internet: www.fda.gov/cder/guidance/index.htm