Der Inhalt dieses Dokuments dient lediglich als Beispiel, die aktuellste Version senden wir gerne zu: info@coachrom.com. The content of this document is an example only. You may request the latest version at: info@coachrom.com

CHROM-2403

Chromogenic Substrate for Plasmin, 25 mg

Ref. 5D-30802

For Research Only. Not for Use in Diagnostic Procedures.
For *in vitro* Use Only.

Description: A chromogenic substrate for the amidolytic assay of Plasmin or Plasminogen-SK.

Formula: pyro-Glu-Phe-Lys-pNA.HCl ($C_{26}H_{32}N_6O_6$.HCl)

Experimental Molecular weight: 561.0

Composition: Enzymatically digestible substrate
K_m: 0.35 mM (Plasmin)
k_{cat}: 60^{S-1} (Plasmin)

Purity, via RP-HPLC: ≥ 95%

Solubility: Up to 25 mg/mL in water

Assay Conditions:

Enzyme activity is determined by measuring the increase in absorbance of the free chromophore (pNA) generated per unit time at a wavelength of 405 nm. At excess substrate concentrations, the rate at which the absorbance increases due to the amount of chromophore released is linearly related to enzyme concentration. Measurement can be made either through acid quenching of the reaction (end-point method), or through use of a kinetic recording spectrophotometer (initial-rate-method).

Presentation:

Ref. 5D-30802 glass vial containing 25 mg (approx. 45 µmoles/vial) of lyophilized, digestible substrate.

Storage and Stability:

Lyophilized substrate may be stored in the dark at 2°-8°C up to the expiration date stated on the vial. Protect from moisture by allowing vial to reach room temperature prior to opening.

Reconstituted substrate may be stored for 1 week at room temperature or 3 months at 2°-8°C. Do not freeze.

Warnings and Precautions:

Hazard Statements:
H315 Causes skin irritation
H319 Causes serious eye irritation
H335 May cause respiratory irritation
Precautionary statements:
P261 Avoid breathing dust
P264 Wash thoroughly after handling
P280 Wear protective gloves/eye protection
P337 + P313 If eye irritation persists: Get medical advice/attention
P403 + P233 Store in a well-ventilated place. Keep container tightly closed.

CoaChrom Diagnostica GmbH, Hauptstrasse 5, A-2344 Maria Enzersdorf, Austria | info@coachrom.com
www.coachrom.com | Tel: +43-1-2362221 or 0800-24 66 330 | Fax: +43-1-236 222111 or 0800-24 66 333