

Der Inhalt dieses Dokuments dient lediglich als Beispiel, die aktuellste Version senden wir Ihnen gerne zu. Bitte anfordern unter: info@coachrom.com. The content of this document is an example only. You may request the latest version at: info@coachrom.com.



BIOPHEN™ Rivaroxaban Calibrator Low

REF 226001

CALI CALII CALIII 4 vials x 1 mL

BIOPHEN™ Rivaroxaban Plasma Calibrator

REF 222701

CAL1 CAL2 CAL3 4 vials x 1 mL



Sales and Support:
CoaChrom Diagnostica GmbH
www.coachrom.com | info@coachrom.com
Tel: +43-1-236 222 1 | Fax: +43-1-236 222 111
Toll-free contact for Germany:
Tel: 0800-24 66 33-0 | Fax: 0800-24 66 33-3

English, revision: 07-2023

INTENDED USE:

For calibration of Rivaroxaban assays, using a quantitative automated method.
This device of *in vitro* diagnostic use is intended for professional use in the laboratory.

SUMMARY AND EXPLANATION:

Technical:

These calibrators are used to establish the calibration curve for anti-Xa chromogenic assays of Rivaroxaban in plasma (BIOPHEN™ DiXal, BIOPHEN™ Heparin LRT, low range / standard range).

Clinical:

Though Rivaroxaban monitoring is not needed in treated patients, measurement in human plasma may be of use in certain cases, in particular in the event of emergency surgery or of suspected overdosage (bleeding risk).

REAGENTS:

CALI Lyophilized human plasma without Rivaroxaban.

CALII Lyophilized human plasma containing approximately 50 ng/mL of Rivaroxaban.

CALIII Lyophilized human plasma containing approximately 100 ng/mL of Rivaroxaban.

CAL1 Lyophilized human plasma without Rivaroxaban.

CAL2 Lyophilized human plasma containing approximately 250 ng/mL of Rivaroxaban.

CAL3 Lyophilized human plasma containing approximately 500 ng/mL of Rivaroxaban.

Calibrator plasmas contain stabilizing agents.

The calibrator concentrations may vary slightly from one batch to another. For the assay, see the exact values indicated on the flyer provided with the kit used.

The product is classified as non-hazardous and is not subject to labeling according to EC Regulation No. 1272/2008 [CLP].

WARNINGS AND PRECAUTIONS:

- Some reagents provided in these kits contain materials of human origin. Whenever human plasma is required for the preparation of these reagents, approved methods are used to test the plasma for the antibodies to HIV 1, HIV 2 and HCV, and for hepatitis B surface antigen, and results are found to be negative. However, no test method can offer complete assurance that infectious agents are absent. Therefore, users of reagents of these types must exercise extreme care in full compliance with safety precautions in the manipulation of these biological materials as if they were infectious.
- Waste should be disposed of in accordance with applicable local regulations.
- Any serious incident that has occurred in relation to the device shall be reported to the manufacturer and the competent authority of the Member State in which the user and/or the patient is established.
- Summary of Safety and Performance (SSP) is available in the European database on medical devices (see Eudamed public website: <https://ec.europa.eu/tools/eudamed>) or on request to HYPHEN BioMed).

REAGENT PREPARATION:

Gently remove the freeze-drying stopper, to avoid any product loss when opening the vial.

CALI **CALII** **CALIII** **CAL1** **CAL2** **CAL3** Reconstitute the contents of each vial with exactly 1 mL of distilled water.

Shake vigorously until complete dissolution while avoiding formation of foam and load it directly on the analyzer following Application Guide instruction.

This plasmatic reagent can be more or less turbid after reconstitution. This turbidity is mainly due to plasma lipids that, after freeze-drying, become "less" soluble and may form a slight deposit. If necessary, let each vial stabilize 10 minutes at room temperature and shake before use.

STORAGE AND STABILITY:

Unopened reagents should be stored at 2-8°C in their original packaging. Under these conditions, they can be used until the expiry date printed on the kit.

CALI **CALII** **CALIII** **CAL1** **CAL2** **CAL3** Reagent stability after reconstitution, free from any contamination or evaporation, and stored closed, is of:

- 7 days at 2-8°C.
- 60 days frozen at -20°C or less*
- Stability on board of the analyzer: see the specific Application Guide.

*Thaw only once, as rapidly as possible at 37°C and use immediately.

REAGENTS AND MATERIALS REQUIRED BUT NOT PROVIDED:

- Laboratory material.

TRACEABILITY:

Lot to lot variability measured on 3 lots is: %CV ≤ 10%.

Calibrators are traceable to European Pharmacopoeia (Ph. Eur.) Certified Reference Standard for Rivaroxaban.

Certificate of traceability and uncertainty is available on the HYPHEN BioMed website:

Uncertainty			
CALI	± 0.0 ng/mL	CAL1	± 0.0 ng/mL
CALII	± 2.4 ng/mL	CAL2	± 15 ng/mL
CALIII	± 4.2 ng/mL	CAL3	± 36 ng/mL

QUALITY CONTROL:

For calibration of Rivaroxaban assays by anti-Xa methods (low range or standard range), with BIOPHEN™ DiXal (221030) and BIOPHEN™ Heparin LRT kits (221011/221013/221015).

The target values are determined from multi-reagent and multi-instrument tests.

The use of quality controls serves to validate method compliance, along with between-series assay homogeneity for a given batch of reagents.

Include the quality controls with each series, as per good laboratory practice, in order to validate the test.

A new calibration curve should be established, preferably for each test series, and at least for each new reagent batch, or after analyzer maintenance, or when the measured quality control values fall outside the acceptance range for the method.

LIMITATIONS:

- If the calibrators are used under measurement conditions other than those validated by HYPHEN BioMed, the test results may vary. The laboratory is responsible for validating the use of these calibrators in its own analytical system.
- Any reagent presenting no limpid appearance or showing signs of contamination must be rejected.

REFERENCES:

- Perzborn E, *et al.* *In vitro* and *in vivo* studies of the novel antithrombotic agent BAY 59-7939-an oral, direct Factor Xa inhibitor. J Thromb Haemost, 2005.
- Kubitza D, *et al.* Safety, pharmacodynamics, and pharmacokinetics of single doses of BAY 59-7939, an oral, direct factor Xa inhibitor. Clin Pharmacol Ther, 2005.
- Mueck W, *et al.* Population pharmacokinetics and pharmacodynamics of once- and twice-daily rivaroxaban for the prevention of venous thromboembolism in patients undergoing total hip replacement. Thromb Haemost 2008.
- Lang D, *et al.* Metabolism and excretion of rivaroxaban - an oral, direct Factor Xa inhibitor - in rats, dogs and humans. Drug Metab Dispos. 2009.
- Rohde G. Determination of rivaroxaban – a novel, oral, direct Factor Xa inhibitor- in human plasma by High-performance liquid chromatography- tandem mass spectrometry. J. Chromatogr. 2008.

e-IFU (other languages) are available on www.hyphen-biomed.com.

For customer support or Application Guides, please contact your local provider or distributor (see www.hyphen-biomed.com).

Changes compared to the previous version.

The following symbols may appear on the product labeling:

REF	Catalogue number	LOT	Batch code	IVD	In-vitro diagnostic medical device
Rx	Numerical < x> identification of reagent		See instructions for use	WHO STD	WHO standard code
	Temperature limitation		Manufacturer		Use by YYYY-MM-DD
CE XXXX	CE marking of conformity with notified body ID number.		Reconstitution volume	CONTENTS	Contents
Cx	Numerical < x> identification of control	I-MA	See instructions in Method Application guide	CONTAINS	Contains
EXP	Expiration date		Contains sufficient for <n> tests	UNIT	Measurement unit
TARGET VALUE	Target Value		Keep away from sunlight and heat	CALx	Numerical < x> identification of calibrator
UDI	Unique Device Identifier		Contains biological material of animal origin		Contains human blood or plasma derivatives
DANGER	Danger	WARNING	Warning	UK CA	UKCA marking of conformity
	Biological risks	ACCEPTANCE RANGE	Acceptance range		