

ZYMUTEST Rat - PAI-1 (Antigen)

RK001A

Antigen PAI-1 (rat)

(Complete one-step ELISA kit for the assay of rat-PAI-1)

FOR RESEARCH USE ONLY.

NOT FOR USE IN DIAGNOSTIC PROCEDURES.



Sales and Support:

CoaChrom Diagnostica GmbH

www.coachrom.com | info@coachrom.com

Tel: +43-1-236 222 1 | Fax: +43-1-236 222 111

Toll-free contact for Germany:

Tel: 0800-24 66 33-0 | Fax: 0800-24 66 33-3

Last revision: 24/06/2014

INTENDED USE:

The ZYMUTEST-rat PAI-1 Antigen kit is a one-step enzyme immuno-assay for measuring rat-PAI-1 (Plasminogen-Activator-Inhibitor-1) in rat plasma, or in any fluid where rat PAI-1 can be present. **This kit is for research use only and should not be used for patient diagnosis or treatment.**

ASSAY PRINCIPLE:

ZYMUTEST-rat-PAI-1 Antigen is a one-step, two site ELISA designed with a pair of monoclonal antibodies specific for rat-PAI-1, and reacting with 2 different and complementary epitopes.

First, the immunoconjugate, which is a monoclonal antibody specific for rat-PAI-1 coupled to Horse Radish Peroxidase (HRP), is introduced into the microwells coated with another monoclonal antibody specific for rat-PAI-1. Then, the diluted tested sample is immediately introduced, and the immunological reaction starts. When present, rat-PAI-1 binds onto the monoclonal antibody coated solid phase through one epitope, and fixes the second monoclonal antibody coupled to HRP by another epitope. Following a washing step, the peroxidase substrate, 3,3',5,5' - Tetramethylbenzidine (TMB), is introduced and a blue colour develops. When the reaction is stopped with Sulfuric Acid, a yellow colour is obtained. The amount of colour developed is directly proportional to the concentration of rat-PA-1 in the tested sample.

TEST SAMPLE:

- Rat plasma collected on trisodium citrate or Na₂ EDTA anticoagulant.
- Any biological fluid where rat-PAI-1 Antigen must be measured.

REAGENTS:

1. **COAT:** Micro ELISA plate, containing 12 strips of 8 wells, coated with a monoclonal antibody specific for rat-PAI-1, then stabilised; the plate is packed in an aluminium pouch hermetically sealed in presence of a desiccant.
2. **SD:** 2 vials containing 50 ml of **F-Sample Diluent**, ready to use.
3. **Std:** 3 vials of **rat-PAI-1 Standard (recombinant)**, lyophilised. Each vial, when restored with 2 ml of **F-Sample diluent (SD)**, allows obtaining the calibrator Solution at "C" ng/ml (about 20 ng/ml). The exact Rat-PAI-1 Ag concentration is indicated on the flyer provided in the kit.
4. **CI:** 1 vial containing 1 ml of lyophilised **rat PAI-1 Control I (High)**.
5. **CL:** 1 vial containing 1 ml of lyophilised **rat PAI-1 Control II (Low)**.

Note: The rat PAI-1: Ag concentrations and acceptancy ranges for controls can vary from lot to lot, and are indicated on the flyer provided in the kit.

6. **IC:** 3 vials of **Anti-(rat)-PAI-1 HRP immunoconjugate**, a mouse monoclonal antibody coupled to HRP, lyophilised.
7. **CD:** 1 vial of 25 ml of **Conjugate Diluent**, ready to use.
8. **WS:** 1 vial of 50 ml of 20 fold concentrated **Wash Solution**.
9. **TMB:** 1 vial of 25 ml peroxidase substrate: 3,3',5,5' - **Tetramethylbenzidine** containing hydrogen peroxide.
10. **SA:** 1 vial of 6 ml of **0.45M Sulfuric acid**.

Note: Use only components from a same kit lot number. Do not mix components from kits with different lots when running the assay.

REAGENTS AND EQUIPMENT REQUIRED BUT NOT PROVIDED:

- **8-channel** or repeating **pipette** allowing dispensing 5-300 µl.
- **1-channel pipettes** at variable volumes from 0 to 20 µl, 20 to 200 µl and 100 to 1000 µl.
- **Micro ELISA plate** washing equipment, and shaker.
- Micro ELISA plate **reader** with a wavelength set up at 450 nm.
- Distilled water.

REAGENTS PREPARATION, STORAGE AND STABILITY:

In their original packaging box, before use, when stored at 2-8°C, the unopened reagents are stable until the expiration date printed on the box.

1. **Micro ELISA plate:** open the aluminium pouch and take off the required amounts of 8 well strips for the test series. When out of the pouch, the strips must be used within 30 minutes. Unused strips can be stored for up to **4 weeks at 2-8°C** in their original aluminium pouch, in presence of the desiccant, hermetically closed and protected from any moisture, and stored in the provided microplate storage bag (minigrip).
2. **F-Sample Diluent:** It is ready to use. When open, it can be used for **4 weeks**, stored at **2-8 °C**, and provided that any bacterial contamination is avoided during use.
3. **Rat-PAI-1 standard:** restore each vial with 2 ml of F-Sample Diluent. This solution is stable for at least **8 hours** at room temperature, and for **24 hours at 2-8°C**.
4. **Rat PAI-1 Control I (high):** restore with 1 ml of F-Sample Diluent (FSD).
5. **Rat PAI-1 Control II (low):** restore with 1 ml of F-Sample Diluent (FSD).

Note: when restored, controls I and II are stable for **8 hours** at room temperature, **24 hours at 2-8°C** or **2 months** frozen at **-20°C** or below.

Warning: Plasma controls I and II (4&5) and standard (3) contain BSA. Any product of biological origin must be handled with all the required cautions, as being potentially infectious.

6. **Anti-(rat)-PAI-1 Immunoconjugate:** each vial must be restored with 4 ml of **conjugate diluent**. Let the pellet to be completely dissolved before use, and shake the vial gently in order to homogenize the content. The restored conjugate is stable for at least 24 hours at room temperature or for at least **4 weeks at 2-8°C**.
7. **Conjugate Diluent:** It is ready to use. When open, it can be used for **4 weeks**, stored at **2-8 °C**, and provided that any bacterial contamination is avoided during use.
8. **Wash Solution:** Incubate the vial for **15-30** minutes in a water bath, at **37°C**, until complete dissolution of solids. Shake the vial and dilute the amount required **1:20** in distilled water (the 50 ml contained in the vial allow preparing 1 liter of wash solution). The wash solution must be stored at 2-8°C in its original vial and used within **4 weeks** following opening. The diluted wash solution must be used within 7 days, when protected from any contamination.
9. **TMB substrate:** It is ready to use. When open, it can be used for **4 weeks**, stored at 2-8°C, and provided that any bacterial contamination is avoided during use.
10. **0.45M Sulfuric acid:** It is ready to use.

Cautions: Sulfuric acid, although diluted to 0.45M is caustic. As for any similar chemical, handle Sulfuric acid with great care. Avoid any skin and eye contact. Wear protection glasses and gloves when handling.

Note: Bring the kit at room temperature, at least 30 min. before use. Store the unused reagents at 2-8°C.

The stability studies at 30°C show that the reagents can be shipped at room temperature without damage.

PROCEDURE:

Specimen collection:

Blood plasma (9 vol.) must be collected on 0.109M citrate anticoagulant (1 vol.); plasma supernatant is decanted following a 20 min. centrifugation at 2,500 g; citrated plasma should be tested within 4 hours or stored frozen at -20°C or below for up to 6 months, and thawed for 15 min. at 37°C just before use. Thawed specimen must be tested within 4 hours. EDTA collected rat plasma may also be used.

Calibration:

Using rat PAI-1 standard provided in the kit, (with a Rat-PAI-1Ag concentration "C" ng/ml indicated, for each lot of reagents, on the flyer provided in the kit), prepare the following standard solutions:

Rat-PAI-1 concentration ng/ml	C	C/2	C/4	C/10	C/20	0 ng/ml
Vol. of standard at "C" ng/ml	1 ml	0.5 ml	0.25 ml	0.1 ml	0.05 ml	0 ml
Vol. of F-Sample Diluent	0 ml	0.5 ml	0.75 ml	0.9 ml	0.95 ml	1 ml

Mix gently for a complete homogenisation.

The standard dilutions are stable for at least **4 hours** at room temperature.

Tested plasma or sample or controls:

The sample must be tested diluted **two fold (1:2)** in the F-Sample Diluent. For expected PAI-1 concentrations > **C or 20 ng/ml**, plasma or samples can be diluted **1:5**, or **1:10**, or **1:20**.

Plasma Controls I and II must be tested diluted **two fold (1:2)**, with F-Sample Diluent.

Assay procedure:

Remove the required number of strips from the aluminium pouch, for the series of measures to be performed. Then put the strips in the frame provided. In the different wells of the micro ELISA plate introduce the reagents and perform the various assay steps as indicated on the following table:

Reagent	Volume	Procedure
Conjugate anti-(rat)-PAI-1-HRP. (Restored with 4 ml of Conjugate Diluent)	100 µl	Introduce the Anti-(rat)-PAI-1- HRP immunoconjugate in the micro ELISA plate wells
Rat-PAI-1 Standard or tested sample or controls or F- Sample Diluent (blank)	100 µl	Introduce immediately the standard solutions or the tested samples in the corresponding micro ELISA plate well (a)
Mix gently on a plate shaker or manually and incubate for 2 hours at room temperature (18-25°C)(b)		
Wash Solution (20 fold diluted in distilled water)	300 µl	Proceed to 5 successive washings using the washing instrument. (c)
TMB/H ₂ O ₂ Substrate	200 µl	Immediately after the washing, introduce the substrate into the wells. Note:The substrate distribution, row by row, must be accurate and at exact time intervals (c,d).
Incubate for exactly 5 minutes at room temperature (18-25 °C) (b)		
0.45 M Sulfuric Acid (SA)	50 µl	Following exactly the same time intervals than for the addition of substrate, stop the colour development by introducing the 0.45M sulfuric acid (d).
Wait for 10 minutes in order to allow the colour to stabilize and measure absorbance at 450 nm (A450) , within the following hours. Subtract the blank value (e).		

Note:

- Distribute calibrators, controls and tested specimen as rapidly as possible, in order to obtain an homogeneous immunological kinetics for antigen binding. A too long delay between the distribution of the first and the last wells may induce an influence of immunological kinetics and produce wrong results.
- Avoid letting the plate in the bright sunlight during incubations and more particularly during colour development. A micro-ELISA plate shaker can be used. An incubation temperature of 18-25°C must be respected. Results are affected by a too high (>25°C) or too low (<18°C) temperature, and measured A450 are then too high or too low. It has to be considered when analyzing the results. In the same way, if a microplate shaker is used, it should be used only at the beginning of each step (sample introduction, immunoconjugate introduction, stop solution introduction), for 1 to 2 minutes, in order to obtain a good homogeneity. A450 values generated in the assay are significantly increased if shaking is used throughout the incubation steps.
- Never let the plates empty between the addition of the reagents or following the washing step. The next reagent must be added within 3 minutes, in order to prevent the plate from drying, which could damage the immobilized components. If necessary, keep the plate filled with Wash Solution and empty it just before the introduction of the next reagent. The washing instrument must be adjusted in order to wash the plates gently, and to avoid a too drastic emptying, which could lower plate reactivity.
- For addition of the TMB substrate, the time interval between each row must be accurate and exactly determined. It must be the same when stopping the reaction.
- For bichromatic readings, a reference wavelength at 690 nm or at 620 nm can be used.

Two-step procedure:

If necessary, the assay can be sensitized by using a two-step procedure. In this case, the conjugate (IC) must be restored with **7.5 ml** of Conjugate Diluent (CD). The other reagents are restored as for the one-step method. The calibration curve is **two fold (1:2)**

diluted respectively to the one step method (i.e. into each well introduce **100 µl** of F-Sample diluent (SD) and **100 µl** of the standard concentration range). The plasma is diluted **two fold (1:2)** directly into the microwell, by introducing **100 µl** of F-Sample Diluent and **100 µl** of plasma (practically, introduce, first, 100 µl of SD into each one of the microwells. Then introduce 100 µl of the rat-PAI-1 standards (Std), ranging from 0 to "C" (about 20 ng/ml), or 100 µl of undiluted rat plasma). The actual dynamic range then goes from **0 to C/2 (about 10 ng/ml)** for the two-fold diluted standards. The assay is performed using a **2 hours** incubation step for rat-PAI-1 binding, followed by a washing step, the introduction of immunoconjugate (IC) (**200 µl/well**), a new **2 hour** incubation step, another washing step, and finally, the colour development is initiated by the introduction of **TMB/H₂O₂** (200 µl/well) substrate, and stopped after 5 min. with **50 µl/well** of SA. The conditions for washing and the cautions are the same than for the one-step method. The measured rat-PAI-1 concentrations must be multiplied by **2** for correcting for the **two fold (1:2)** dilution factor.

RESULTS:

- On a linear graph paper plot the **rat PAI-1: Ag concentrations (ng/ml)** on abscissa and the corresponding absorbances (**A450**) on ordinates.
- Users must construct their own calibration curve, obtained using their standard dilutions (see model on the flyer). From the curve obtained, deduce the rat PAI-1: Ag concentration for the tested dilution. For obtaining the rat PAI-1: Ag concentration in the tested sample, the value read on the calibration curve must be **multiplied by the dilution factor (i.e. 2, 5, 10, 20.....)**
- For controls I and II, the concentrations measured must be multiplied by **2**.
- Alternatively, an ELISA software (i.e. Dynex, Biolise, etc...) can be used for the calculation of concentrations.

The results obtained should be for research purposes only and not used for patient diagnosis or treatment

EXPECTED RANGE:

- PAI-1:Ag concentration in normal rat plasma is of **1.8 ± 0.9 ng/ml** (ref. 3).
- In lysed platelet rich rat plasma the PAI-1:Ag concentration is of **6 ± 1 ng/ml**.
- In plasma from lipopolysaccharide treated rats, the PAI-1:Ag concentration is dramatically increased (**200 to > 1,000 ng/ml**).

ASSAY REACTIVITY:

ZYMUTEST rat-PAI-1 Antigen is reactive with latent, active or inactive rat-PAI-1 as well as with rat-tPA-PAI-1 complexes. The assay then measures all the rat-PAI-1 forms present in the tested sample.