

# Certificate of Analysis

## Fibrinogen Deficient Plasma

**Article no.:** FG-FDP

**Lot / Exp.:** xxxxxx / xxxx-xx

**Description:** Pooled normal citrated human plasma depleted of Fibrinogen under controlled conditions, using purified human thrombin. Plasma contains 20 mM HEPES and stabilizers.

**Presentation:** Frozen Fibrinogen deficient plasma.

**Preparation:** Thaw vials in 37°C water bath for 5 minutes.

**Storage:** Frozen deficiency plasma should be stored below -60 until expiration date stated on vial label. Once thawed, plasma is stable for 4 hours at 2-8°C in original vial.

**Specifications:**

PARAMETER	RESULT
APTT (CEPHEN APTT)	>180 sec
PT (THROMBOPHEN)	>120 sec
FACTOR II (ACTIVITY)	116 %
FACTOR V (ACTIVITY)	56 %
FACTOR VIII (ACTIVITY)	88 %
FIBRINOGEN (ANTIGEN BY ELISA)	<0.1 g/L

Human plasma was tested and found negative by FDA accepted methods for Anti-HIV 1/2, Anti-HTLV I & II, HBsAg, Anti-HCV, Syphilis, HBC Ab, HIV-1 p24 Ag or HIV-1 RNA, HCV RNA and HBV RNA. Donors are screened for CJD (Creutzfeld-Jakob Disease). Nevertheless, it should be handled by personnel trained in the proper procedures for handling potential viral contaminants.

For Research Use Only. Not for Use in Diagnostic Procedures.