

Certificate of Analysis

Fibrinogen Deficient Plasma

Article no.: FG-FDP

Lot / Exp.: xxxxxx / xxxx-xx

Description: Pooled normal citrated human plasma depleted of Fibrinogen under controlled conditions, using purified human thrombin. Plasma contains 20 mM HEPES and stabilizers.

Presentation: Frozen Fibrinogen deficient plasma.

Preparation: Thaw vials in 37°C water bath for 5 minutes.

Storage: Frozen deficiency plasma should be stored at -60 to -80°C until expiration date stated on vial label. Once thawed, plasma is stable for 4 hours at 2-8°C in original vial.

Specifications:

PARAMETER	RESULT
APTT (CEPHEN APTT)	>180 sec
PT (THROMBOPHEN)	>120 sec
FACTOR II (ACTIVITY)	95%
FACTOR V (ACTIVITY)	32%
FACTOR VIII (ACTIVITY)	101%
FIBRINOGEN (ANTIGEN BY ELISA)	0.059 g/L

Human plasma was tested and found negative by FDA accepted methods for Anti-HIV 1/2, Anti-HTLV I & II, HBsAg, Anti-HCV, Syphilis, HBC Ab, HIV-1 p24 Ag or HIV-1 RNA, HCV RNA and HBV RNA. Donors are screened for CJD (Creutzfeld-Jakob Disease). Nevertheless, it should be handled by personnel trained in the proper procedures for handling potential viral contaminants.

For Research Use Only. Not for Use in Diagnostic Procedures.